

# OPTIME – time and frequency dissemination system based on fiber optical network – local repository project

W. Adamowicz<sup>4</sup>, A. Binczewski<sup>5</sup>, W. Bogacki<sup>5</sup>, A. Czubla<sup>3</sup>, P. Dunst<sup>2</sup>, J. Igalson<sup>4</sup>, J. Kołodziej<sup>1</sup>, P. Krehlik<sup>1</sup>, D. Lemański<sup>2</sup>, M. Lipiński<sup>1</sup>, J. Nawrocki<sup>2</sup>, P. Nogaś<sup>2</sup>, P. Ostapowicz<sup>5</sup>, T. Pawszak<sup>4</sup>, J. Pieczerak<sup>4</sup>, M. Stroiński<sup>5</sup>, Ł. Śliwczyński<sup>1</sup>, K. Turza<sup>5</sup>

<sup>1</sup> AGH University of Science and Technology (AGH), Krakow, Poland

<sup>2</sup> Astrogdynamic Observatory (AOS), Borowiec, Poland

<sup>3</sup> Central Office of Measures (GUM), Warsaw, Poland

<sup>4</sup> Orange Polska S.A. (TPSA), Warsaw, Poland

<sup>5</sup> Poznan Supercomputing and Networking Center (PSNC), Poznan, Poland

Email: mlipinsk@agh.edu.pl, nawrocki@cbk.poznan.pl, a.czubla@gum.gov.pl,

Janusz.Pieczerak@orange.com, wojbor@man.poznan.pl,

The document describes a current stage of creating a high precision dissemination system for time and frequency reference signals - OPTIME. The OPTIME system is based on fiber optical network in Poland. The first part of document focuses on an operational 420-km long link between time/frequency laboratories: GUM in Warsaw and AOS in Borowiec. The comparison of UTC(PL) and UTC(AOS) which are provided by laboratories will be referred (Fig 1.)




Fig. 1: Time variance - short term stability of the comparisons UTC(AOS) - UTC(PL) is very high, for averaging times of 10-20 s it is limited by the precession of the used counter (Stanford SR-620)




Fig. 2: Local time & frequency repository

The second part of document presents all parts of OPTIME system. In particular a local time /frequency repository (Fig 2.) will be shown. This repository is a one of the key elements of the OPTIME system. It acts as a proxy system between reference laboratories and end users. Moreover it also provides a backup signals at the time of connection failure between repository and reference laboratories. The monitoring and control system will be also described with a hardware and software components. The last part shows a current status of building a local repository in PSNC - Poznan. Acknowledgement: Project OPTIME (no. PBS1/A3/13/2012) is co-funded by The National Centre for Research and Development – Poland.