

# A Compact 461-nm Laser Source for $\text{Sr}^+$ Trapped-Ion Experiments

Thomas Fordell, Thomas Lindvall, Anders E. Wallin, Mikko Merimaa

Centre for Metrology and Accreditation (MIKES), Espoo, Finland

Email: thomas.fordell@mikes.fi

Strontium can be efficiently photoionized in a two-step process<sup>1</sup> that requires light at 461 nm (excitation) and 405 nm (ionization via a broad auto-ionizing state). The latter step is easily achieved using cheap, free-running laser diodes; however, the first step is more demanding since the laser frequency should, preferably, be correct to within a few hundred MHz. Here we describe a compact, stabilized source of 461 nm radiation appropriate for a future transportable  $\text{Sr}^+$  single-ion clock.

The setup is shown in Fig. 1. A 921-nm, 100-mW DBR laser diode (Photodigm PH921.7DBR) is frequency doubled in a fiber-coupled PPLN waveguide<sup>2</sup>. After beam shaping by an anamorphic prism pair and optical isolation, about 50 mW is available for fiber coupling and second-harmonic generation. For low input powers, the conversion efficiency of the waveguide is 230%/W, yielding >2mW of light at 461 nm. A chromatic beam splitter separates the second harmonic beam from the fundamental, the latter of which is sent to a frequency comb (optional stabilization). About 90% of the second harmonic is sent to the ion trap, while the remaining 10% is used for frequency stabilization, that is, for measuring the

absorption in a beam of Sr atoms emitted from a dispenser in a small vacuum cell<sup>3</sup>. To maximize the lifetime of the dispenser, balanced lock-in detection is used; however, the sensitivity of this scheme is compromised by the sensitivity of the PPLN waveguide to frequency modulation; the resulting amplitude modulation cannot fully be corrected by the balanced detector. In a future upgrade, the vacuum cell should be equipped with another view port (top side) to be used for background-free imaging of the scattered light using a low cost CCD camera. In the beam direction, the scattering is position dependent due to the Doppler effect, i.e., the frequency is mapped into a position enabling frequency stabilization via spatial imaging.

This work was supported by the Academy of Finland (project 138894) and the European Metrology Research Program (EMRP). The EMRP is jointly funded by the EMRP participating countries within EURAMET and the EU. TF acknowledges financial support from the European Commission (Marie Curie Integration Grant PCIG10-GA-2011-304084).




Fig. 1: Experimental setup. LD: laser diode; ISO: isolator; WG: PPLN waveguide; DBS: dichroic beam splitter; BS: beam splitter. The physical size of the setup is approximately 30 cm x 45 cm x 10 cm.

<sup>1</sup> M. Brownnutt *et al.*, “Controlled photoionization loading of  $^{88}\text{Sr}^+$  for precision ion-trap experiments”, *Appl. Phys. B*, vol. 87, p. 411-415, 2007.

<sup>2</sup> D. Akamatsu *et al.*, “A compact light source at 461 nm using a periodically poled  $\text{LiNbO}_3$  waveguide for strontium magneto-optical trapping”, *Opt. Express*, vol. 19, p. 2046-2051, 2013.

<sup>3</sup> E. M. Bridge *et al.*, “A vapor cell based on dispensers for laser spectroscopy”, *Rev. Sci. Instr.*, vol. 80, 013101, 2009.